

# CALTRONIX

## Certificate of Calibration

CALTRONIX certifies that this instrument has been calibrated to published specifications using measurement standards traceable to the National Institute of Standards and Technology (NIST). CALTRONIX's calibration system meets or exceeds the requirements of MIL-STD-45662A, ANSI/NCSL Z540-1, ISO/IEC Guide 25 and ISO/IEC 17025.

CALTRONIX is accredited to ISO/IEC 17025 by ACLASS, Assured Calibration and Laboratory Accreditation Select Services, Certificate #'s AC-1182 & AC-1183.

Customer :	UNIVERSITY OF BUFFALO	Test Uncertainty Ratio:> or =	4 :1
	WHITE ROAD	Procedure :	CAL-8080
	BUFFALO, NY 14260	Interval :	12 Months
Description :	VOLT CALIBRATOR	Date Received :	07/19/2007
Manufacturer :	FLUKE	Date Calibrated:	07/25/2007
Model Number :	715	Date Due :	07/25/2008
Serial Number :	9005337	Relative Humidity < or =	50 %
Company ID :		Temperature	74 F +/- 3.0 F
R.O. Number :	7K959		IN TOLERANCE

Uncertainty budgets have been calculated at a 95% confidence level using a coverage factor of k=2.

Calibration Facility : 100 Town Centre Drive <> Rochester, New York 14623

Asset	Manufacturer	Model No.	Description	Calcd	Due
3015	KEITHLEY	2001	HIGH PRECISION MULTIMETER	06/15/2007	09/15/2007
9042	FLUKE	5520A	CALIBRATOR	10/13/2006	10/13/2007

NIST	Cap	Sandia 49426	Volts	Fluke JJA	Dimen	821/270003
Numbers	Ind	811/264817	Res	811/269310	Mass	822/264749
	Freq	USNO (GPS)	Temp	255156		



JAMES R. COYLE  
Calibration Technician

Michael A. Dent 07/25/2007  
Director of Quality

CALTRONIX Inc. <> 100 Town Centre Drive <> Rochester, New York 14623 <> 585 359-3780 <> Fax 585 334-0064

This certificate may not be altered or reproduced, except in full, without written approval of CALTRONIX.

C A L I B R A T I O N   D A T A   S H E E T

CALTRONIX, Inc.  
 100 Town Centre Drive  
 Rochester, NY 14623  
 (585) 359-3780   FAX (585) 334-0064

R.O. Number	7K959	Calibrated By	30	Date Calibrated	07/25/07
Manufacturer	FLUKE	Company	UNIVERSITY OF BUFFALO		
Model	715	Serial	9005337		
Description	VOLT CALIBRATOR	ID No.			

Function/Parameter Tested	Lo Limit	As Found	Hi Limit	As Left
<b>DC VOLTAGE MEASURE</b>				
0.0000mV	-0.02	0.00	0.02	0.00
50.0000mV	49.98	50.01	50.02	50.01
150.0000mV	149.96	150.04	150.04	150.04
0.0000V	-0.002	0.000	0.002	0.000
10.0000V	9.997	10.001	10.003	10.001
20.0000V	19.996	20.003	20.004	20.003
<b>DC CURRENT MEASURE</b>				
4.0000mA	3.998	4.001	4.002	4.001
12.0000mA	11.996	12.003	12.004	12.003
24.0000mA	23.993	24.006	24.007	24.006
<b>DC VOLTAGE SOURCE</b>				
0.00mV	-0.02	0.01	0.02	0.01
50.00mV	49.98	49.99	50.02	49.99
150.00mV	149.96	149.96	150.04	149.96
0.000V	-0.002	.001	0.002	.001
5.000V	4.998	5.000	5.002	5.000
15.000V	14.997	14.998	15.003	14.998
<b>DC CURRENT SOURCE</b>				
4.000mA	3.9972	3.9974	4.0028	3.9974
12.000mA	11.9956	11.9966	12.0044	11.9966
24.000mA	23.9932	23.9944	24.0068	23.9944